



RAS FORM 5

RADIATION SURVEY REPORT (SEALED SOURCE LEAK TEST)

1. Authorized User: _____
2. Leak test performed by (please print): _____ Date: _____
3. Department: _____
4. Building and room number: _____
5. Instrument used for wipe sample counting: _____

Sample #	Source	Location	DPM ¹	microcuries ²	Blank

Sealed sources must be checked for leaks using wipe tests every six months.

1. DPM = [Scintillation counter sample results in CPM] / [Scintillation Counter Efficiency]
2. 1 Microcurie = 2.2 x 10⁶ DPM